

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2182

Name: RITTER 1  
Sterilizer ID: V2178198  
07/26/2022 04:06 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 154.4 F 0.0  
2:00 161.6 F 2.8  
4:00 182.2 F 7.3  
6:00 204.1 F 12.7  
8:00 220.2 F 7.5  
10:00 235.4 F 13.0  
12:00 235.8 F 10.1  
14:00 238.1 F 8.2  
16:00 248.9 F 15.0  
18:00 260.9 F 21.7  
20:00 272.0 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.7  
1:00 272.6 F 29.2  
2:00 271.8 F 28.8  
3:00 272.7 F 29.3  
4:00 271.7 F 28.6

Min 271.6 F 28.6  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.8 F 0.0  
5:00 196.9 F 0.0  
10:00 210.4 F 0.0  
15:00 215.3 F 0.0  
20:00 216.7 F 0.0  
25:00 217.5 F 0.0  
30:00 217.8 F 0.0  
Drying Complete

FILLING: 1:32  
HEATING: 20:32  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:59:17  
CYCLE COMPLETE

07/26/2022 05:05 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_