
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000281
15:43 01/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:54
PRESSURIZING 3:05
STERILIZING 4:36
136.0 C 228kPa 4:36
136.4 C 233kPa 5:06
136.4 C 233kPa 5:36
136.4 C 233kPa 6:06
136.4 C 233kPa 6:36
136.4 C 233kPa 7:06
136.4 C 233kPa 7:36
136.4 C 233kPa 8:06
136.4 C 233kPa 8:06
VENTING 8:07
AIR DRYING 9:11
CYCLE COMPLETE 24:11

Digital Signature #
4ED9C8503B605DBF
