
Midmark M11 - v1.0.5
Total Cycles: 964

Name: RITTER2
Sterilizer ID: V2313228
04/06/2022 03:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.3 F 0.0
2:00 153.9 F 0.1
4:00 181.8 F 0.8
6:00 203.3 F 5.8
8:00 220.9 F 11.2
10:00 228.4 F 8.0
12:00 239.0 F 11.6
14:00 243.2 F 12.9
16:00 243.8 F 12.6
18:00 257.8 F 20.1
20:00 269.8 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.0
1:00 272.6 F 29.4
2:00 272.2 F 29.5
3:00 272.6 F 29.3
4:00 272.1 F 29.4

Min 271.7 F 28.9
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.9 F 0.0
5:00 193.7 F 0.0
10:00 209.0 F 0.0
15:00 217.1 F 0.0
20:00 222.5 F 0.0
25:00 222.1 F 0.0
30:00 222.7 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 20:55
STERILIZING: 4:00
VENTING: 2:59
DRYING: 30:00
TOTAL CYCLE: 00:58:56
CYCLE COMPLETE

04/07/2022 06:59 AM
ACCEPTED: X
REJECTED:
Operator: U02