
Midmark M11 - v1.0.5
Total Cycles: 1575

Name: RITTER2
Sterilizer ID: V2313228
10/06/2022 12:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.0 F 0.0
2:00 174.3 F 0.1
4:00 186.1 F 2.3
6:00 205.8 F 7.5
8:00 222.7 F 13.1
10:00 232.5 F 9.3
12:00 238.0 F 9.8
14:00 245.2 F 12.9
16:00 245.4 F 13.2
18:00 258.0 F 20.1
20:00 269.4 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.8 F 29.5
2:00 271.8 F 28.7
3:00 272.3 F 29.5
4:00 272.3 F 29.2

Min 271.4 F 28.6
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.2 F 0.0
5:00 235.8 F 0.0
10:00 230.3 F 0.0
15:00 228.3 F 0.0
20:00 230.2 F 0.0
25:00 227.6 F 0.0
30:00 229.6 F 0.0
Drying Complete

FILLING: 1:28
HEATING: 21:02
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:43
CYCLE COMPLETE

10/06/2022 01:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____