

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002402  
11:16 20/12/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:51  
PRESSURIZING 1:58  
STERILIZING 2:59  
136.5 C 223kPa 2:59  
136.3 C 224kPa 3:29  
136.2 C 224kPa 3:59  
136.4 C 223kPa 4:29  
136.1 C 224kPa 4:59  
136.2 C 224kPa 5:29  
136.4 C 225kPa 5:59  
136.4 C 225kPa 6:29  
136.4 C 225kPa 6:29  
VENTING 6:29  
AIR DRYING 7:19  
CYCLE COMPLETE 37:19  
-----

Digital Signature #  
437D773C6386E0F0  
-----