

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4180

Name: RITTER 1  
Sterilizer ID: V2178198  
01/04/2024 12:16 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 150.1 F 0.0  
2:00 152.7 F 2.7  
4:00 169.6 F 6.3  
6:00 190.7 F 11.2  
8:00 222.0 F 5.4  
10:00 231.7 F 11.1  
12:00 230.9 F 7.3  
14:00 245.9 F 13.4  
16:00 243.1 F 11.6  
18:00 255.7 F 18.2  
20:00 267.0 F 25.2  
22:00 271.3 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.5  
1:00 271.6 F 28.4  
2:00 272.1 F 28.6  
3:00 272.8 F 28.8  
4:00 272.5 F 29.0

Min 271.2 F 28.4  
Max 272.9 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 173.8 F 0.0  
5:00 194.1 F 0.0  
10:00 194.4 F 0.0  
15:00 195.0 F 0.0  
20:00 192.6 F 0.0  
25:00 189.2 F 0.0  
30:00 186.4 F 0.0  
Drying Complete

FILLING: 1:08  
HEATING: 22:00  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 01:00:28  
CYCLE COMPLETE

01/04/2024 01:16 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_