

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4999

Name: RITTER 1  
Sterilizer ID: V2178198  
10/16/2024 11:41 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 128.9 F 0.0  
2:00 137.4 F 0.1  
4:00 155.2 F 0.3  
6:00 185.1 F 1.4  
8:00 201.2 F 4.5  
10:00 211.8 F 8.1  
12:00 222.7 F 11.8  
14:00 226.3 F 6.3  
16:00 235.9 F 11.1  
18:00 233.1 F 8.0  
20:00 243.0 F 11.9  
22:00 245.9 F 13.0  
24:00 257.5 F 19.2  
26:00 267.9 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.0  
1:00 271.8 F 28.4  
2:00 272.4 F 28.8  
3:00 272.3 F 28.7  
4:00 272.1 F 28.5

Min 271.2 F 27.9  
Max 272.5 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.5 F 0.0  
5:00 212.0 F 0.0  
10:00 191.9 F 0.0  
15:00 184.8 F 0.0  
20:00 181.7 F 0.0  
25:00 182.1 F 0.0  
30:00 205.5 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 27:35  
STERILIZING: 4:00  
VENTING: 3:57  
DRYING: 30:00  
TOTAL CYCLE: 01:06:18  
CYCLE COMPLETE

10/16/2024 12:47 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_