
Midmark M11 - v1.0.5
Total Cycles: 3668

Name: RITTER2
Sterilizer ID: V2313228
06/05/2024 08:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.0 F 0.0
2:00 80.6 F 0.1
4:00 129.6 F 0.5
6:00 111.3 F 2.4
8:00 148.3 F 5.0
10:00 161.5 F 8.2
12:00 186.4 F 11.9
14:00 217.7 F 4.3
16:00 220.1 F 7.8
18:00 227.0 F 11.5
20:00 227.4 F 6.7
22:00 239.4 F 11.7
24:00 236.1 F 9.1
26:00 248.6 F 14.9
28:00 258.9 F 20.6
30:00 268.4 F 26.8
32:00 271.7 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.4
1:00 273.0 F 29.8
2:00 272.1 F 29.9
3:00 272.4 F 29.5
4:00 272.6 F 30.1

Min 271.3 F 29.2
Max 273.3 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.3 F 0.0
5:00 233.8 F 0.0
10:00 225.1 F 0.0
15:00 187.6 F 0.0
20:00 182.2 F 0.0
25:00 176.8 F 0.0
30:00 176.2 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 32:02
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:10:10
CYCLE COMPLETE

06/05/2024 09:35 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____