
Midmark M11 - v1.0.5
Total Cycles: 1393

Name: RITTER2
Sterilizer ID: V2313228
08/15/2022 02:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.0 F 0.0
2:00 159.4 F 0.1
4:00 184.4 F 2.2
6:00 211.3 F 6.8
8:00 218.0 F 11.9
10:00 226.4 F 8.0
12:00 239.7 F 13.5
14:00 241.1 F 11.8
16:00 240.8 F 11.0
18:00 254.2 F 17.8
20:00 265.7 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 272.4 F 29.5
2:00 271.6 F 28.7
3:00 272.0 F 29.5
4:00 272.3 F 29.2

Min 271.2 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.7 F 0.0
5:00 236.5 F 0.0
10:00 220.2 F 0.0
15:00 215.3 F 0.0
20:00 203.6 F 0.0
25:00 201.6 F 0.0
30:00 201.8 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 21:43
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:03
CYCLE COMPLETE

08/15/2022 03:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____