
Midmark M11 - v1.0.5
Total Cycles: 1485

Name: RITTER2
Sterilizer ID: V2313228
09/13/2022 12:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.3 F 0.0
2:00 166.3 F 3.0
4:00 177.7 F 7.4
6:00 199.7 F 12.6
8:00 221.2 F 7.1
10:00 236.3 F 12.4
12:00 235.1 F 10.0
14:00 236.4 F 8.6
16:00 249.4 F 15.7
18:00 261.6 F 22.7
20:00 272.6 F 30.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.7 F 30.1
2:00 272.1 F 29.4
3:00 271.6 F 29.5
4:00 272.8 F 30.0

Min 271.3 F 29.1
Max 273.0 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.9 F 0.0
5:00 227.4 F 0.0
10:00 232.4 F 0.0
15:00 226.1 F 0.0
20:00 228.6 F 0.0
25:00 224.6 F 0.0
30:00 222.9 F 0.0
Drying Complete

FILLING: 1:52
HEATING: 20:24
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:26
CYCLE COMPLETE

09/13/2022 01:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____