

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1371

Name: RITTER 1  
Sterilizer ID: V2178198  
11/15/2021 01:32 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 187.1 F 0.0  
2:00 175.7 F 3.1  
4:00 195.8 F 7.5  
6:00 216.2 F 13.1  
8:00 230.4 F 8.2  
10:00 241.1 F 12.8  
12:00 240.2 F 11.2  
14:00 237.9 F 9.7  
16:00 252.0 F 16.5  
18:00 264.0 F 23.3  
20:00 271.1 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.4  
1:00 272.9 F 29.3  
2:00 272.0 F 29.1  
3:00 272.0 F 28.6  
4:00 272.6 F 29.0  
  
Min 271.0 F 28.2  
Max 273.0 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.8 F 0.0  
5:00 232.6 F 0.0  
10:00 208.3 F 0.0  
15:00 206.5 F 0.0  
20:00 211.0 F 0.0  
25:00 223.9 F 0.0  
30:00 232.7 F 0.0  
Drying Complete

FILLING: 2:23  
HEATING: 20:04  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:59:42  
CYCLE COMPLETE

11/15/2021 02:36 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02