
Midmark M11 - v1.0.5
Total Cycles: 2377

Name: RITTER2
Sterilizer ID: V2313228
05/16/2023 01:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.1 F 0.0
2:00 188.0 F 4.0
4:00 206.4 F 9.6
6:00 222.4 F 6.9
8:00 234.4 F 11.8
10:00 236.6 F 10.2
12:00 236.3 F 9.4
14:00 251.8 F 16.9
16:00 264.4 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.2
1:00 272.2 F 29.4
2:00 272.3 F 29.9
3:00 271.9 F 29.2
4:00 272.6 F 29.9

Min 271.4 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.3 F 0.0
5:00 216.8 F 0.0
10:00 230.5 F 0.0
15:00 235.8 F 0.0
20:00 228.7 F 0.0
25:00 225.2 F 0.0
30:00 215.9 F 0.0
Drying Complete

FILLING: 2:28
HEATING: 17:49
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 00:57:36
CYCLE COMPLETE

05/16/2023 02:50 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____