
Midmark M11 - v1.0.5
Total Cycles: 4338

Name: RITTER2
Sterilizer ID: V2313228
02/06/2025 01:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 105.0 F 0.0
2:00 106.3 F 0.6
4:00 110.6 F 2.3
6:00 129.2 F 5.0
8:00 159.0 F 8.6
10:00 182.1 F 12.4
12:00 207.5 F 4.5
14:00 221.1 F 8.1
16:00 230.7 F 12.0
18:00 230.5 F 8.3
20:00 242.1 F 13.3
22:00 240.8 F 11.6
24:00 252.3 F 17.4
26:00 262.8 F 23.5
28:00 272.2 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.4
1:00 272.2 F 30.3
2:00 272.4 F 30.5
3:00 272.6 F 30.3
4:00 272.6 F 30.2

Min 270.9 F 29.3
Max 273.2 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.8 F 0.0
5:00 230.4 F 0.0
10:00 221.2 F 0.0
15:00 214.6 F 0.0
20:00 197.8 F 0.0
25:00 206.1 F 0.0
30:00 224.4 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 28:33
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:07:13
CYCLE COMPLETE

02/06/2025 02:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____