
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000988
8:50 16/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:58
PRESSURIZING 3:01
STERILIZING 4:07
136.4 C 224kPa 4:07
136.1 C 221kPa 4:37
136.3 C 224kPa 5:07
136.0 C 219kPa 5:37
136.1 C 222kPa 6:07
136.1 C 222kPa 6:37
136.4 C 222kPa 7:07
136.3 C 220kPa 7:37
136.3 C 220kPa 7:37
VENTING 7:37
AIR DRYING 8:31
CYCLE COMPLETE 8:43

Digital Signature #
DAA155DB749B3DDC
