
Midmark M11 - v1.0.5
Total Cycles: 4471

Name: RITTER2
Sterilizer ID: V2313228
03/28/2025 12:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 120.6 F 0.0
2:00 123.2 F 0.9
4:00 126.6 F 3.1
6:00 142.1 F 6.2
8:00 162.1 F 9.9
10:00 186.8 F 9.0
12:00 213.1 F 6.3
14:00 226.9 F 10.1
16:00 232.3 F 9.6
18:00 235.3 F 11.2
20:00 235.4 F 9.0
22:00 247.5 F 14.9
24:00 257.9 F 20.6
26:00 267.3 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 30.0
1:00 271.3 F 30.4
2:00 272.2 F 31.0
3:00 273.1 F 31.4
4:00 272.5 F 30.7

Min 270.9 F 29.9
Max 274.1 F 31.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.7 F 0.0
5:00 184.2 F 0.0
10:00 201.8 F 0.0
15:00 203.8 F 0.0
20:00 202.4 F 0.0
25:00 205.1 F 0.0
30:00 206.5 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 27:46
STERILIZING: 4:00
VENTING: 3:57
DRYING: 30:00
TOTAL CYCLE: 01:06:36
CYCLE COMPLETE

03/28/2025 01:31 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____