
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001802
8:39 26/06/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:38
PRESSURIZING 2:45
STERILIZING 3:40
136.0 C 220kPa 3:40
136.1 C 222kPa 4:10
136.2 C 224kPa 4:40
136.4 C 226kPa 5:10
136.3 C 224kPa 5:40
136.3 C 223kPa 6:10
136.4 C 224kPa 6:40
136.3 C 223kPa 7:10
136.3 C 223kPa 7:10
VENTING 7:10
AIR DRYING 8:02
CYCLE COMPLETE 23:36

Digital Signature #
CB7AA4E0BD8C2178
