
Midmark M11 - v1.0.5
Total Cycles: 1065

Name: RITTER2
Sterilizer ID: V2313228
05/03/2022 01:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.6 F 0.0
2:00 173.5 F 3.2
4:00 187.1 F 7.9
6:00 207.6 F 13.3
8:00 222.0 F 8.0
10:00 235.9 F 13.5
12:00 238.6 F 11.5
14:00 239.0 F 10.5
16:00 252.8 F 17.5
18:00 264.7 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.6 F 30.0
2:00 272.3 F 29.5
3:00 271.4 F 29.1
4:00 272.2 F 30.0

Min 271.4 F 29.0
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.9 F 0.0
5:00 214.0 F 0.0
10:00 221.2 F 0.0
15:00 220.0 F 0.0
20:00 189.6 F 0.0
25:00 184.9 F 0.0
30:00 177.2 F 0.0
Drying Complete

FILLING: 1:44
HEATING: 19:54
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:52
CYCLE COMPLETE

05/03/2022 02:39 PM
ACCEPTED: X
REJECTED:
Operator: U02