
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000551
13:46 30/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:59
PRESSURIZING 3:03
STERILIZING 3:55
136.0 C 222kPa 3:55
136.2 C 224kPa 4:25
136.4 C 225kPa 4:55
136.4 C 223kPa 5:25
136.4 C 225kPa 5:55
136.4 C 224kPa 6:25
136.4 C 225kPa 6:55
136.4 C 225kPa 7:25
136.4 C 225kPa 7:25
VENTING 7:25
AIR DRYING 8:16
CYCLE COMPLETE 8:19

Digital Signature #
31B79E31D4B66634
