

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3363

Name: RITTER2  
Sterilizer ID: V2313228  
02/07/2024 11:31 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 153.6 F 0.0  
2:00 156.2 F 1.1  
4:00 174.9 F 4.2  
6:00 197.0 F 8.2  
8:00 210.9 F 12.6  
10:00 221.3 F 6.5  
12:00 232.3 F 11.0  
14:00 231.6 F 7.8  
16:00 244.4 F 12.8  
18:00 244.6 F 12.7  
20:00 256.6 F 19.0  
22:00 267.3 F 25.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.7  
1:00 272.6 F 29.1  
2:00 271.4 F 28.7  
3:00 272.4 F 29.2  
4:00 271.4 F 28.5

Min 271.3 F 28.5  
Max 272.6 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.4 F 0.0  
5:00 206.6 F 0.0  
10:00 191.7 F 0.0  
15:00 190.4 F 0.0  
20:00 187.2 F 0.0  
25:00 183.7 F 0.0  
30:00 181.6 F 0.0  
Drying Complete

FILLING: 1:18  
HEATING: 23:29  
STERILIZING: 4:00  
VENTING: 3:37  
DRYING: 30:00  
TOTAL CYCLE: 01:02:23  
CYCLE COMPLETE

02/07/2024 12:33 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_