
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.9uS / 0.5ppm
CYCLE NUMBER 001514
13:35 10/20/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:10
PRESSURIZING 7:02
STERILIZING 8:34
136.2 C 220kPa 8:34
136.4 C 222kPa 9:04
136.2 C 222kPa 9:34
136.1 C 219kPa 10:04
136.4 C 220kPa 10:34
136.4 C 222kPa 11:04
136.3 C 222kPa 11:35
136.3 C 221kPa 12:04
VENTING 12:04
AIR DRYING 13:09
CYCLE COMPLETE 36:59

Digital Signature #
4CB731EAADB37060
