

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1830

Name: RITTER 1  
Sterilizer ID: V2178198  
04/18/2022 07:04 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.5 F 0.0  
2:00 73.3 F 0.0  
4:00 92.8 F 0.0  
6:00 127.1 F 0.0  
8:00 153.6 F 2.3  
10:00 176.3 F 5.7  
12:00 202.1 F 9.0  
14:00 223.9 F 12.5  
16:00 222.0 F 4.9  
18:00 232.3 F 8.7  
20:00 242.1 F 12.8  
22:00 237.3 F 9.2  
24:00 241.6 F 10.4  
26:00 245.8 F 12.6  
28:00 257.1 F 18.7  
30:00 267.7 F 25.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.0  
1:00 272.6 F 28.6  
2:00 271.2 F 27.8  
3:00 272.6 F 28.8  
4:00 271.9 F 28.1  
  
Min 271.1 F 27.8  
Max 272.7 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.1 F 0.0  
5:00 223.8 F 0.0  
10:00 200.7 F 0.0  
15:00 207.6 F 0.0  
20:00 222.2 F 0.0  
25:00 224.2 F 0.0  
30:00 226.5 F 0.0  
Drying Complete

FILLING: 0:45  
HEATING: 31:24  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:09:19  
CYCLE COMPLETE

04/18/2022 08:13 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_