
Midmark M11 - v1.0.5
Total Cycles: 2537

Name: RITTER2
Sterilizer ID: V2313228
06/20/2023 10:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.5 F 0.0
2:00 192.0 F 3.3
4:00 208.5 F 8.5
6:00 226.4 F 9.6
8:00 233.4 F 10.5
10:00 235.2 F 7.8
12:00 241.8 F 11.1
14:00 246.2 F 13.8
16:00 258.9 F 20.7
18:00 270.2 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.7
1:00 272.4 F 29.7
2:00 272.4 F 29.2
3:00 271.3 F 28.6
4:00 271.3 F 29.1

Min 271.0 F 28.6
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.1 F 0.0
5:00 216.9 F 0.0
10:00 240.8 F 0.0
15:00 226.1 F 0.0
20:00 218.4 F 0.0
25:00 219.9 F 0.0
30:00 229.4 F 0.0
Drying Complete

FILLING: 2:08
HEATING: 18:56
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:14
CYCLE COMPLETE

06/20/2023 11:41 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____