

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3292

Name: RITTER2  
Sterilizer ID: V2313228  
01/12/2024 10:09 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.7 F 0.0  
2:00 72.3 F 0.0  
4:00 77.3 F 0.0  
6:00 93.6 F 0.6  
8:00 132.4 F 2.6  
10:00 161.4 F 5.6  
12:00 182.8 F 8.7  
14:00 199.2 F 12.0  
16:00 210.6 F 4.3  
18:00 220.2 F 7.7  
20:00 229.2 F 11.5  
22:00 229.0 F 7.2  
24:00 240.7 F 12.4  
26:00 239.0 F 10.5  
28:00 250.8 F 16.3  
30:00 261.2 F 22.3  
32:00 271.1 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 29.1  
1:00 272.1 F 29.7  
2:00 272.4 F 29.4  
3:00 271.7 F 29.4  
4:00 272.9 F 29.8  
  
Min 271.5 F 29.0  
Max 272.9 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.8 F 0.0  
5:00 214.2 F 0.0  
10:00 219.7 F 0.0  
15:00 215.1 F 0.0  
20:00 213.4 F 0.0  
25:00 214.5 F 0.0  
30:00 217.3 F 0.0  
  
Drying Complete

FILLING: 0:50  
HEATING: 32:37  
STERILIZING: 4:00  
VENTING: 3:31  
DRYING: 30:00  
TOTAL CYCLE: 01:10:58  
CYCLE COMPLETE

01/12/2024 11:19 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_