

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 002652  
7:59 07/04/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:30  
PRESSURIZING 2:34  
STERILIZING 3:48  
136.0 C 220kPa 3:48  
136.3 C 223kPa 4:18  
136.4 C 224kPa 4:48  
136.1 C 219kPa 5:18  
136.2 C 221kPa 5:48  
136.4 C 222kPa 6:18  
136.2 C 221kPa 6:48  
136.2 C 223kPa 7:18  
136.2 C 223kPa 7:18  
VENTING 7:18  
AIR DRYING 8:13  
CYCLE COMPLETE 8:16

-----  
Digital Signature #  
BB8136818CE4A887  
-----