
Midmark M11 - v1.0.0
Total Cycles: 1090

Name: RITTER 1
Sterilizer ID: V2178198
08/18/2021 07:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.4 F 0.0
2:00 72.8 F 0.0
4:00 79.8 F 0.0
6:00 114.8 F 0.0
8:00 148.4 F 0.1
10:00 169.1 F 0.2
12:00 194.1 F 2.3
14:00 211.9 F 5.4
16:00 221.8 F 8.9
18:00 231.1 F 12.5
20:00 230.3 F 7.6
22:00 241.3 F 12.4
24:00 238.4 F 9.7
26:00 237.9 F 8.8
28:00 249.6 F 14.8
30:00 261.4 F 21.4
32:00 272.0 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.0
1:00 272.3 F 28.7
2:00 271.8 F 28.0
3:00 272.3 F 28.7
4:00 272.1 F 28.2

Min 271.3 F 27.8
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.9 F 0.0
5:00 202.9 F 0.0
10:00 192.0 F 0.0
15:00 192.7 F 0.0
20:00 207.8 F 0.0
25:00 218.8 F 0.0
30:00 235.9 F 0.0

Drying Complete

FILLING: 0:58
HEATING: 32:24
STERILIZING: 4:00
VENTING: 3:01
DRYING: 30:00
TOTAL CYCLE: 01:10:23
CYCLE COMPLETE

08/18/2021 08:27 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____