
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.8uS / 0.5ppm
CYCLE NUMBER 000297
11:09 13/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:18
PRESSURIZING 4:21
STERILIZING 5:36
136.0 C 230kPa 5:36
136.4 C 233kPa 6:06
136.3 C 231kPa 6:36
136.4 C 233kPa 7:06
136.4 C 233kPa 7:36
136.4 C 233kPa 8:06
136.3 C 232kPa 8:36
136.3 C 229kPa 9:06
136.3 C 229kPa 9:06
VENTING 9:06
AIR DRYING 9:59
CYCLE COMPLETE 10:02

Digital Signature #
11EB2948C7B7018C
