

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1950

Name: RITTER 1  
Sterilizer ID: V2178198  
05/25/2022 10:17 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 164.8 F 0.0  
2:00 162.8 F 3.4  
4:00 176.4 F 7.2  
6:00 197.7 F 11.9  
8:00 216.9 F 5.7  
10:00 229.1 F 10.7  
12:00 229.6 F 6.8  
14:00 243.4 F 12.8  
16:00 241.0 F 10.8  
18:00 253.7 F 17.2  
20:00 264.9 F 23.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.3  
1:00 272.0 F 28.9  
2:00 272.3 F 28.7  
3:00 271.3 F 28.2  
4:00 272.5 F 29.2

Min 271.2 F 28.1  
Max 272.5 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.6 F 0.0  
5:00 211.5 F 0.0  
10:00 226.4 F 0.0  
15:00 236.5 F 0.0  
20:00 228.6 F 0.0  
25:00 225.8 F 0.0  
30:00 224.8 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 21:44  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:59:50  
CYCLE COMPLETE

05/25/2022 11:16 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_