

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5129

Name: RITTER 1  
Sterilizer ID: V2178198  
11/22/2024 03:48 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 149.7 F 0.0  
2:00 149.5 F 0.0  
4:00 175.6 F 0.3  
6:00 202.9 F 1.2  
8:00 218.3 F 5.6  
10:00 228.9 F 10.0  
12:00 234.5 F 8.8  
14:00 240.8 F 11.9  
16:00 238.7 F 10.1  
18:00 235.9 F 8.6  
20:00 251.2 F 15.9  
22:00 263.3 F 22.9  
24:00 271.7 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 272.6 F 28.9  
2:00 271.5 F 28.2  
3:00 272.6 F 29.0  
4:00 271.8 F 28.3

Min 271.4 F 28.1  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.7 F 0.0  
5:00 190.3 F 0.0  
10:00 190.5 F 0.0  
15:00 194.6 F 0.0  
20:00 213.0 F 0.0  
25:00 219.8 F 0.0  
30:00 223.5 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 24:07  
STERILIZING: 4:00  
VENTING: 3:26  
DRYING: 30:00  
TOTAL CYCLE: 01:02:24  
CYCLE COMPLETE

11/22/2024 04:50 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_