

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4939

Name: RITTER 1  
Sterilizer ID: V2178198  
09/25/2024 03:16 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 108.6 F 0.0  
2:00 112.8 F 0.0  
4:00 138.0 F 0.1  
6:00 167.3 F 0.3  
8:00 188.7 F 2.6  
10:00 202.1 F 5.7  
12:00 208.9 F 9.3  
14:00 222.3 F 13.2  
16:00 225.4 F 8.0  
18:00 237.1 F 12.7  
20:00 237.7 F 10.0  
22:00 237.4 F 9.0  
24:00 248.5 F 14.4  
26:00 259.8 F 20.6  
28:00 269.8 F 27.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.2  
1:00 271.9 F 28.8  
2:00 272.5 F 29.0  
3:00 272.2 F 28.6  
4:00 271.5 F 28.3

Min 271.1 F 28.1  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.1 F 0.0  
5:00 206.4 F 0.0  
10:00 186.6 F 0.0  
15:00 187.7 F 0.0  
20:00 202.0 F 0.0  
25:00 220.2 F 0.0  
30:00 227.8 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 29:08  
STERILIZING: 4:00  
VENTING: 3:47  
DRYING: 30:00  
TOTAL CYCLE: 01:07:46  
CYCLE COMPLETE

09/25/2024 04:23 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_