

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1212

Name: RITTER2  
Sterilizer ID: V2313228  
06/22/2022 11:07 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 182.6 F 0.0  
2:00 177.9 F 3.6  
4:00 189.8 F 8.0  
6:00 209.6 F 12.9  
8:00 223.2 F 7.9  
10:00 234.3 F 13.2  
12:00 238.3 F 11.0  
14:00 237.9 F 9.9  
16:00 251.6 F 16.7  
18:00 263.2 F 23.5  
20:00 272.0 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.1  
1:00 272.9 F 29.8  
2:00 271.8 F 29.2  
3:00 271.3 F 29.2  
4:00 272.4 F 29.8  
  
Min 271.2 F 28.9  
Max 272.9 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.1 F 0.0  
5:00 223.7 F 0.0  
10:00 226.5 F 0.0  
15:00 229.1 F 0.0  
20:00 227.6 F 0.0  
25:00 228.4 F 0.0  
30:00 223.7 F 0.0  
Drying Complete

FILLING: 1:50  
HEATING: 20:11  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:13  
CYCLE COMPLETE

06/22/2022 12:06 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_