
Midmark M11 - v1.0.5
Total Cycles: 4388

Name: RITTER2
Sterilizer ID: V2313228
02/27/2025 04:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 116.9 F 0.0
2:00 115.0 F 0.9
4:00 122.6 F 2.8
6:00 163.2 F 5.8
8:00 171.3 F 9.5
10:00 188.8 F 13.5
12:00 216.2 F 6.1
14:00 226.7 F 10.1
16:00 233.5 F 9.4
18:00 237.8 F 11.2
20:00 235.7 F 9.2
22:00 248.4 F 15.3
24:00 259.2 F 21.4
26:00 269.3 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 30.0
1:00 272.4 F 30.0
2:00 272.6 F 30.3
3:00 272.2 F 30.6
4:00 271.7 F 30.0

Min 271.7 F 29.9
Max 272.9 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.7 F 0.0
5:00 213.2 F 0.0
10:00 209.6 F 0.0
15:00 204.8 F 0.0
20:00 200.6 F 0.0
25:00 193.9 F 0.0
30:00 191.1 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 27:51
STERILIZING: 4:00
VENTING: 3:53
DRYING: 30:00
TOTAL CYCLE: 01:06:44
CYCLE COMPLETE

02/27/2025 05:14 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____