

-----  
Midmark M11 - v1.0.5  
Total Cycles: 785

Name: RITTER2  
Sterilizer ID: V2313228  
02/10/2022 04:40 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 147.9 F 0.0  
2:00 155.2 F 2.4  
4:00 184.7 F 6.7  
6:00 209.0 F 11.9  
8:00 225.0 F 6.8  
10:00 234.8 F 12.1  
12:00 235.9 F 9.7  
14:00 236.5 F 9.1  
16:00 248.9 F 15.3  
18:00 261.4 F 22.3  
20:00 272.4 F 29.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.1  
1:00 272.7 F 29.9  
2:00 271.7 F 29.1  
3:00 271.8 F 29.7  
4:00 272.6 F 29.8

Min 271.1 F 29.0  
Max 272.9 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.2 F 0.0  
5:00 241.2 F 0.0  
10:00 216.5 F 0.0  
15:00 216.0 F 0.0  
20:00 219.7 F 0.0  
25:00 220.1 F 0.0  
30:00 236.7 F 0.0  
Drying Complete

FILLING: 2:18  
HEATING: 20:28  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:59:56  
CYCLE COMPLETE

02/10/2022 05:48 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02