

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 007664  
12:34 11/07/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:33  
PRESSURIZING 2:39  
STERILIZING 3:41  
136.5 C 222kPa 3:41  
136.4 C 225kPa 4:11  
136.2 C 221kPa 4:41  
136.2 C 220kPa 5:11  
136.4 C 220kPa 5:41  
136.0 C 222kPa 6:11  
136.4 C 224kPa 6:41  
136.4 C 224kPa 7:11  
136.4 C 224kPa 7:11  
VENTING 7:11  
AIR DRYING 8:03  
CYCLE COMPLETE 20:47

-----  
Digital Signature #  
61F180DF1DE6FCCC  
-----