
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001311
12:32 20/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:12
PRESSURIZING 3:18
STERILIZING 4:23
136.5 C 220kPa 4:23
136.3 C 225kPa 4:53
136.1 C 223kPa 5:23
136.4 C 224kPa 5:53
136.2 C 223kPa 6:23
136.1 C 222kPa 6:53
136.3 C 224kPa 7:23
136.2 C 223kPa 7:53
136.2 C 223kPa 7:53
VENTING 7:53
AIR DRYING 8:45
CYCLE COMPLETE 23:29

Digital Signature #
EBD56192A7D718C4
