

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
2.3uS / 1.4ppm  
CYCLE NUMBER 000068  
10:30 17/07/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:03  
PRESSURIZING 3:09  
STERILIZING 4:38  
136.4 C 225kPa 4:38  
136.4 C 225kPa 5:08  
136.3 C 224kPa 5:38  
136.4 C 225kPa 6:08  
136.3 C 224kPa 6:38  
136.4 C 225kPa 7:08  
136.2 C 223kPa 7:38  
136.2 C 225kPa 8:08  
136.2 C 225kPa 8:08  
VENTING 8:08  
AIR DRYING 9:00  
CYCLE COMPLETE 24:00  
-----

Digital Signature #  
2651027AB8EA4945  
-----