
Midmark M11 - v1.0.0
Total Cycles: 981

Name: RITTER 1
Sterilizer ID: V2178198
07/08/2021 07:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.1 F 0.0
2:00 184.8 F 2.8
4:00 206.5 F 7.2
6:00 219.6 F 13.5
8:00 229.7 F 8.7
10:00 238.8 F 10.2
12:00 243.2 F 12.0
14:00 240.4 F 10.4
16:00 254.6 F 17.5
18:00 266.5 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.1
1:00 272.2 F 28.3
2:00 272.5 F 28.8
3:00 271.2 F 28.2
4:00 271.9 F 28.4

Min 271.0 F 27.9
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.2 F 0.0
5:00 241.5 F 0.0
10:00 213.7 F 0.0
15:00 199.1 F 0.0
20:00 193.9 F 0.0
25:00 191.6 F 0.0
30:00 238.9 F 0.0
Drying Complete

FILLING: 2:18
HEATING: 19:29
STERILIZING: 4:00
VENTING: 2:55
DRYING: 30:00
TOTAL CYCLE: 00:58:42
CYCLE COMPLETE

07/08/2021 08:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____