

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1509

Name: RITTER2  
Sterilizer ID: V2313228  
09/20/2022 01:12 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 175.2 F 0.0  
2:00 171.1 F 3.0  
4:00 182.8 F 7.6  
6:00 205.7 F 13.4  
8:00 223.2 F 7.9  
10:00 237.3 F 12.4  
12:00 239.6 F 11.5  
14:00 239.3 F 10.5  
16:00 253.3 F 17.5  
18:00 265.2 F 24.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.0  
1:00 272.6 F 29.6  
2:00 271.5 F 28.9  
3:00 272.7 F 29.8  
4:00 272.1 F 29.2  
  
Min 271.5 F 28.9  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.6 F 0.0  
5:00 215.0 F 0.0  
10:00 224.5 F 0.0  
15:00 228.7 F 0.0  
20:00 228.1 F 0.0  
25:00 224.5 F 0.0  
30:00 221.4 F 0.0  
Drying Complete

FILLING: 1:45  
HEATING: 19:45  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:58:40  
CYCLE COMPLETE

09/20/2022 02:10 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_