

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
1.1uS / 0.7ppm  
CYCLE NUMBER 000311  
12:06 20/11/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:13  
PRESSURIZING 3:17  
STERILIZING 4:27  
136.0 C 221kPa 4:27  
136.4 C 233kPa 4:57  
136.4 C 232kPa 5:27  
136.2 C 232kPa 5:57  
136.2 C 231kPa 6:27  
136.2 C 232kPa 6:57  
136.2 C 232kPa 7:27  
136.3 C 232kPa 7:57  
136.3 C 232kPa 7:57  
VENTING 7:58  
AIR DRYING 8:50  
CYCLE COMPLETE 8:57  
-----

Digital Signature #  
671EBBD7DB3859AF  
-----