
Midmark M11 - v1.0.5
Total Cycles: 1383

Name: RITTER2
Sterilizer ID: V2313228
08/09/2022 10:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 204.0 F 0.0
2:00 199.4 F 2.1
4:00 195.6 F 7.3
6:00 215.0 F 12.8
8:00 227.4 F 8.7
10:00 237.7 F 10.6
12:00 242.2 F 12.8
14:00 242.4 F 12.0
16:00 255.4 F 18.8
18:00 266.7 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.2
1:00 272.6 F 29.6
2:00 271.8 F 29.4
3:00 272.8 F 29.6
4:00 271.6 F 29.1

Min 271.6 F 29.0
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.4 F 0.0
5:00 216.7 F 0.0
10:00 235.9 F 0.0
15:00 229.9 F 0.0
20:00 224.8 F 0.0
25:00 222.9 F 0.0
30:00 219.6 F 0.0

Drying Complete

FILLING: 1:40
HEATING: 19:31
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:20
CYCLE COMPLETE

08/09/2022 11:35 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____