

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3858

Name: RITTER2  
Sterilizer ID: V2313228  
08/14/2024 12:12 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 143.7 F 0.0  
2:00 152.1 F 2.6  
4:00 173.2 F 5.9  
6:00 193.3 F 9.9  
8:00 209.5 F 7.5  
10:00 221.0 F 7.0  
12:00 232.3 F 11.5  
14:00 232.6 F 8.3  
16:00 243.0 F 12.2  
18:00 244.6 F 13.0  
20:00 256.5 F 19.2  
22:00 267.0 F 25.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.9  
1:00 272.7 F 29.6  
2:00 271.7 F 28.8  
3:00 272.1 F 29.7  
4:00 272.2 F 29.3

Min 271.3 F 28.7  
Max 272.7 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.9 F 0.0  
5:00 216.1 F 0.0  
10:00 215.1 F 0.0  
15:00 197.3 F 0.0  
20:00 191.0 F 0.0  
25:00 187.7 F 0.0  
30:00 184.4 F 0.0  
Drying Complete

FILLING: 1:15  
HEATING: 23:32  
STERILIZING: 4:00  
VENTING: 3:41  
DRYING: 30:00  
TOTAL CYCLE: 01:02:29  
CYCLE COMPLETE

08/14/2024 01:14 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_