

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2421

Name: RITTER2  
Sterilizer ID: V2313228  
05/26/2023 12:45 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 138.6 F 0.0  
2:00 137.5 F 0.0  
4:00 149.5 F 2.1  
6:00 178.2 F 6.0  
8:00 196.6 F 10.2  
10:00 212.9 F 6.7  
12:00 225.2 F 8.2  
14:00 236.2 F 12.7  
16:00 236.9 F 9.9  
18:00 238.1 F 9.4  
20:00 248.2 F 14.5  
22:00 259.5 F 20.8  
24:00 269.7 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.7  
1:00 272.5 F 29.4  
2:00 271.5 F 28.6  
3:00 272.4 F 29.5  
4:00 272.2 F 29.0

Min 271.4 F 28.6  
Max 272.6 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.8 F 0.0  
5:00 213.9 F 0.0  
10:00 217.6 F 0.0  
15:00 219.2 F 0.0  
20:00 212.8 F 0.0  
25:00 210.9 F 0.0  
30:00 213.3 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 25:01  
STERILIZING: 4:00  
VENTING: 3:23  
DRYING: 30:00  
TOTAL CYCLE: 01:03:20  
CYCLE COMPLETE

05/26/2023 01:48 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_