
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 008395
7:12 11/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:56
PRESSURIZING 2:57
STERILIZING 4:00
136.4 C 221kPa 4:00
136.0 C 221kPa 4:30
136.4 C 221kPa 5:00
136.2 C 219kPa 5:30
136.1 C 221kPa 6:00
CYCLE ABORTED 6:25
STOP BUTTON PRESSED
NOT STERILE
END TIME 6:25

Digital Signature #
B17D1691E0247667
