
Midmark M11 - v1.0.5
Total Cycles: 3240

Name: RITTER2
Sterilizer ID: V2313228
12/20/2023 04:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.7 F 0.0
2:00 166.5 F 3.0
4:00 189.6 F 7.2
6:00 207.3 F 11.8
8:00 226.3 F 6.3
10:00 229.2 F 11.4
12:00 235.1 F 8.6
14:00 241.1 F 10.8
16:00 246.5 F 13.7
18:00 258.5 F 20.1
20:00 269.0 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 272.5 F 29.1
1:00 272.4 F 29.3
2:00 272.8 F 29.5
3:00 272.2 F 29.7
4:00 271.4 F 29.2

Min 271.3 F 29.0
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.8 F 0.0
5:00 223.0 F 0.0
10:00 219.4 F 0.0
15:00 191.8 F 0.0
20:00 188.7 F 0.0
25:00 187.6 F 0.0
30:00 186.8 F 0.0
Drying Complete

FILLING: 1:31
HEATING: 21:39
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:00:43
CYCLE COMPLETE

12/20/2023 05:23 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____