

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4254

Name: RITTER 1  
Sterilizer ID: V2178198  
01/31/2024 03:02 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 167.4 F 0.0  
2:00 168.0 F 2.7  
4:00 186.2 F 6.4  
6:00 205.1 F 11.6  
8:00 223.1 F 5.5  
10:00 229.3 F 10.4  
12:00 232.5 F 7.7  
14:00 244.2 F 12.9  
16:00 242.5 F 11.4  
18:00 254.3 F 17.6  
20:00 265.6 F 24.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.1  
1:00 272.7 F 29.3  
2:00 272.0 F 28.9  
3:00 272.0 F 28.6  
4:00 272.3 F 29.0

Min 270.7 F 28.0  
Max 273.1 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.4 F 0.0  
5:00 245.1 F 0.0  
10:00 238.3 F 0.0  
15:00 210.7 F 0.0  
20:00 209.3 F 0.0  
25:00 227.4 F 0.0  
30:00 205.5 F 0.0  
Drying Complete

FILLING: 1:38  
HEATING: 21:43  
STERILIZING: 4:00  
VENTING: 3:33  
DRYING: 30:00  
TOTAL CYCLE: 01:00:54  
CYCLE COMPLETE

01/31/2024 04:02 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_