
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000391
10:42 25/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:03
PRESSURIZING 2:08
STERILIZING 3:09
136.5 C 233kPa 3:09
136.1 C 231kPa 3:39
136.3 C 233kPa 4:09
136.2 C 233kPa 4:39
136.3 C 233kPa 5:09
136.3 C 233kPa 5:39
136.4 C 231kPa 6:09
136.0 C 231kPa 6:39
136.0 C 231kPa 6:39
VENTING 6:39
AIR DRYING 7:26
CYCLE COMPLETE 7:44

Digital Signature #
F36087D07885729B
