

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4130

Name: RITTER2  
Sterilizer ID: V2313228  
11/18/2024 12:41 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 127.2 F 0.0  
2:00 127.4 F 0.0  
4:00 134.1 F 0.1  
6:00 164.9 F 3.2  
8:00 189.1 F 6.7  
10:00 207.8 F 10.9  
12:00 221.5 F 6.3  
14:00 230.7 F 10.7  
16:00 233.3 F 7.9  
18:00 243.2 F 13.5  
20:00 243.0 F 12.4  
22:00 254.7 F 18.6  
24:00 265.5 F 25.1  
26:00 271.7 F 29.7

STERILIZING  
mm:ss Degrees PSI  
0:00 272.3 F 29.6  
1:00 272.3 F 29.6  
2:00 272.0 F 29.6  
3:00 272.1 F 29.6  
4:00 272.1 F 29.7

Min 271.6 F 29.5  
Max 272.8 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.6 F 0.0  
5:00 205.5 F 0.0  
10:00 202.2 F 0.0  
15:00 191.9 F 0.0  
20:00 188.5 F 0.0  
25:00 196.7 F 0.0  
30:00 197.6 F 0.0  
Drying Complete

FILLING: 0:47  
HEATING: 26:19  
STERILIZING: 4:00  
VENTING: 3:38  
DRYING: 30:00  
TOTAL CYCLE: 01:04:44  
CYCLE COMPLETE

11/18/2024 01:45 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_