
Midmark M11 - v1.0.5
Total Cycles: 4131

Name: RITTER2
Sterilizer ID: V2313228
11/18/2024 03:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 91.6 F 0.0
2:00 93.3 F 0.0
4:00 105.9 F 0.0
6:00 143.6 F 0.1
8:00 173.3 F 1.1
10:00 196.3 F 3.8
12:00 200.1 F 7.2
14:00 211.7 F 10.4
16:00 230.1 F 9.6
18:00 229.3 F 9.1
20:00 240.3 F 13.0
22:00 239.5 F 11.4
24:00 237.6 F 9.9
26:00 249.9 F 15.8
28:00 260.5 F 21.7
30:00 270.2 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.2 F 29.9
2:00 272.6 F 29.5
3:00 271.5 F 29.0
4:00 271.2 F 29.2

Min 271.0 F 28.9
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.3 F 0.0
5:00 206.2 F 0.0
10:00 212.0 F 0.0
15:00 210.7 F 0.0
20:00 208.9 F 0.0
25:00 208.8 F 0.0
30:00 212.8 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 30:59
STERILIZING: 4:00
VENTING: 3:56
DRYING: 30:00
TOTAL CYCLE: 01:09:43
CYCLE COMPLETE

11/18/2024 04:38 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____