

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2686

Name: RITTER 1  
Sterilizer ID: V2178198  
12/15/2022 03:38 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 169.2 F 0.0  
2:00 194.8 F 2.8  
4:00 203.5 F 7.0  
6:00 218.9 F 13.0  
8:00 222.7 F 7.7  
10:00 237.8 F 13.1  
12:00 237.8 F 10.5  
14:00 237.3 F 8.8  
16:00 251.2 F 16.0  
18:00 263.6 F 23.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.5  
1:00 272.6 F 28.9  
2:00 271.7 F 28.6  
3:00 272.6 F 29.0  
4:00 271.6 F 28.4  
  
Min 271.6 F 28.3  
Max 272.8 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.6 F 0.0  
5:00 211.7 F 0.0  
10:00 222.3 F 0.0  
15:00 230.0 F 0.0  
20:00 232.8 F 0.0  
25:00 227.5 F 0.0  
30:00 228.1 F 0.0  
  
Drying Complete

FILLING: 1:44  
HEATING: 19:59  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:58:57  
CYCLE COMPLETE

12/15/2022 04:36 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_