
Midmark M11 - v1.0.0
Total Cycles: 2659

Name: RITTER 1
Sterilizer ID: V2178198
12/08/2022 11:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 209.7 F 0.0
2:00 203.9 F 3.1
4:00 204.1 F 7.9
6:00 218.3 F 11.1
8:00 230.8 F 8.9
10:00 241.0 F 10.4
12:00 242.8 F 12.2
14:00 241.5 F 11.0
16:00 254.9 F 17.9
18:00 266.6 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.6 F 29.0
2:00 271.8 F 28.3
3:00 272.1 F 28.9
4:00 272.5 F 28.9

Min 271.5 F 28.1
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.3 F 0.0
5:00 223.8 F 0.0
10:00 245.3 F 0.0
15:00 238.8 F 0.0
20:00 227.3 F 0.0
25:00 224.8 F 0.0
30:00 220.8 F 0.0

Drying Complete

FILLING: 1:34
HEATING: 19:33
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:15
CYCLE COMPLETE

12/08/2022 12:10 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____