

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
1.0uS / 0.6ppm  
CYCLE NUMBER 001534  
12:35 10/27/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:47  
PRESSURIZING 6:39  
STERILIZING 8:13  
136.1 C 220kPa 8:13  
136.4 C 222kPa 8:44  
136.1 C 220kPa 9:14  
136.1 C 217kPa 9:44  
136.3 C 222kPa 10:14  
136.4 C 220kPa 10:44  
136.0 C 221kPa 11:14  
136.1 C 218kPa 11:43  
VENTING 11:44  
AIR DRYING 12:46  
CYCLE COMPLETE 18:45  
-----

Digital Signature #  
37BBF39CEAF25C9C  
-----